

mumu

SURGICAL MASK

3-PLY SURGICAL MASK WITH EARLOOP

According to Annex B of EN 14683:2019 + AC:2019
Type IIR and "BFE % \geq 98"

PURPOSE OF USE

Face mask can be used for hygienic applications.
Suitable for use by physicians and patients.
Made of non-woven fabric. Air permeable structure
allows easy breathing. Does not contain allergic materials.
3 ply structure. Non-irritating.



Compatible Tire



Filtered and 3-Ply



**U.S. FOOD & DRUG
ADMINISTRATION**



MY TİCARET VE MEDİKAL A.Ş.
Ömerli Mah. General Şükrü Koraltı Cad
No:33 Arnavutköy / İstanbul / TÜRKİYE
Tel: 0212 438 20 64 Faks: 0212 438 20 65
info@mymedikal.com.tr

www.mymedikal.com.tr



MY Medikal



MANAGEMENT SYSTEM
ISO/IEC 17021-1:2015
NAC-002-MS

Certificate of Registration

This is to certify that

Quality Management System

of

MY TİCARET VE MEDİKAL ANONİM ŞİRKETİ

ÖMERLİ MAHALLESİ GENERAL ŞÜKRÜ KORALTI CAD. NO: 33
ARNAVUTKÖY - İSTANBUL / TÜRKİYE

complies with requirements of

ISO 9001:2015

This certificate is valid concerning all activities related to;

PRODUCTION AND SALE OF LATEX POWDERED / POWDER FREE EXAMINATION GLOVES, NITRILE POWDER FREE EXAMINATION GLOVES, STERIL / NON-STERILE SURGICAL GLOVES, STERIL / NON-STERILE SPONGE GAUZE COMPRESS, STERIL / NON-STERIL COMPRESSE ABDOMINALE, STERIL / NON-STERILE COTTON PAD, GAUZE, STERILE / NON-STERILE SURGICAL MASK

LATEKS PUDRALI / PUDRASIZ MUAYENE ELĐİVENİ, NİTRİL PUDRASIZ MUAYENE ELĐİVENİ, STERİL / NON-STERİL CERRAHi ELĐİVEN, STERİL / NON-STERİL SPANÇ GAZ KOMPRES, STERİL / NON-STERİL BATIN KOMPRES, STERİL / NON-STERİL PAMUKLU PED, GAZLI BEZ, STERİL / NON-STERİL CERRAHi MASKE ÜRETİMİ VE SATIŞI

ISO 01 940 1179

Certificate No.

Jun. 5, 2020

Date of this Certificate

Jun. 4, 2021

Certification Expiry Date

May. 28, 2020

Date of Audit

Jun. 5, 2020

Date of Registration

Managing Director / Director



Medicert Uluslararası Ürün Ve Sistem Belgelendirme Ltd. Şti.
Tersane Mah. Cemal Gürsel Cad. No:11/3 Halide Hnm. Apt. Karşıyaka / İzmir
Tel: 0232 327 33 44 www.medicert.com.tr info@medicert.com.tr

* You can query the validity of this certificate by sending an e-mail to info@medicert.com.tr.



Certificate of Registration

This is to certify that

Quality Management System
for Medical Devices

of

MY TİCARET VE MEDİKAL ANONİM ŞİRKETİ

ÖMERLİ MAHALLESİ GENERAL ŞÜKRÜ KORALTI CAD. NO: 33
ARNAVUTKÖY - İSTANBUL / TÜRKİYE

complies with requirements of

ISO 13485:2016

This certificate is valid concerning all activities related to;

SALES OF LATEX POWDERED / POWDER-FREE EXAMINATION GLOVES- NITRILE POWDER-FREE EXAMINATION GLOVES- POWDERED / POWDER-FREE STERILE SURGICAL GLOVES-VINYL POWDERED / POWDER-FREE EXAMINATION GLOVES. PRODUCTION AND SALE OF DISPOSABLE NON-STERILE MASKS.

LATEKS PUDRALI/PUDRASIZ MUAYENE ELDİVENİ- NİTRİL PUDRASIZ MUAYENE ELDİVENİ- PUDRALI/PUDRASIZ STERİL CERRAHİ ELDİVEN-VİNİL PUDRALI/PUDRASIZ MUAYENE ELDİVENİ SATIŞI. TEK KULLANIMLIK NON-STERİL MASKE ÜRETİMİ VE SATIŞI.

ISO 02 836 1179
Certificate No.

Feb. 26, 2020
Date of this Certificate

Feb. 25, 2021
Certification Expiry Date

Feb. 21, 2020
Date of Audit

Feb. 26, 2020
Date of Registration


Managing Director / Director



Medicert Uluslararası Ürün Ve Sistem Belgelendirme Ltd. Şti.
Tersane Mah. Cemal Gürsel Cad. No:11/3 Halide Hnm. Apt. Karşıyaka / İzmir
Tel: 0232 327 33 44 www.medicert.com.tr info@medicert.com.tr

* You can query the validity of this certificate by sending an e-mail to info@medicert.com.tr.



Certificate of Registration 2020

This is to certify that the registration of

MY TICARET VE MEDIKAL A.S
OMERLI MAH. GENERAL SUKRU KORALTI CAD. NO: 33 ARNAVUTKOY,
ISTANBUL, TURKEY - 34555

with U.S. Food and Drug Administration as required by 21 CFR Part 807 is successfully completed by Liberty Management Group Ltd with the information provided by My Ticaret Ve Medikal A.S

Owner/Operator Number	10075681
Date of Registration	June 23, 2020
Date of Expiration	December 31, 2020
US Agent	Liberty Management Group Ltd.
Device Listing Numbers	See Annex
Certificate Number	3006230220

This certificate does not make representations or warranties to any person or entity other than the named certificate holder; it is issued for record keeping purpose only. This certificate does not denote endorsement or approval of certificate holder's facility or product by the U.S. food and Drug Administration. Liberty management Group Ltd. assumes no liability to any person or entity in connection with the foregoing.

The U.S. Food and Drug Administration does not issue a certificate of registration, nor does the U.S. Food and Drug Administration recognize a certificate of registration. Liberty Management Group Ltd. is not affiliated with the U.S. Food and Drug Administration.

LMG LIBERTY
MANAGEMENT
GROUP LTD.

75 Executive Drive, Aurora, Illinois, USA
www.fdahelp.us

A handwritten signature in black ink, appearing to read 'Manoj Zacharias'.

Manoj Zacharias

President

Liberty Management Group LTD.

Dated: June 23, 2020



Certificate of Registration 2020

Annex - Device Listings

Listing Number	Code	Device Name - Proprietary Names
D409537	QKR	Face mask (except N95 respirator) for general public/healthcare personnel per IIE guidance - Mumu Surgical Mask



EC DECLARATION OF CONFORMITY

- Manufacturer's Name:** *MY TİCARET VE MEDİKAL A.Ş.*
- Manufacturer Address:** *Ömerli Mah. General Şükrü Koraltı Cad. No:33
Arnavutköy/İSTANBUL*
- Medical Devices:** *Surgical Mask, 3-ply with earloop
Ref No: MM.NS.LM.01*
- Classification:** *Medical Device Directive-Annex IX, Rule I, Class-I (Type IIR)*
- GMDN Code and Term:** *57794 / Surgical – Medical Respirator*
- Scope of Application:** *All batches supplied to which the Declaration of Conformity
Procedure has been applied.*
- Declaration:** *Conformity of the products has been assessed in accordance
with Annex VII of the Directive. A dossier of technical
documentation, as required by the Directive is available. The
product listed is designed, manufactured and tested in
accordance with the information set out in the dossier. We
declare our products comply with EN 14683:2019+AC:2019 as
Type IIR*
- Verification Certificates:** *Quality Management System- Medical devices
EN ISO13485:2016 Certificate No: ISO 02 836 1179
Quality Management System
EN ISO 9001:2015 Certificate No: ISO 01 940 117*

Standards Applied:

<i>EN ISO 13485</i>	<i>Medical devices-Quality management systems - Requirements for regulatory purposes</i>
<i>EN ISO 9001:2015</i>	<i>Quality management systems</i>
<i>EN ISO 15223-1</i>	<i>Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied</i>
<i>MDD 93/42/EEC</i>	<i>Medical devices directive</i>
<i>EN ISO 1041</i>	<i>Information supplied by the manufacturer of medical devices</i>
<i>EN ISO 14683:2019+AC: 2019</i>	<i>Medical face masks - Requirements and test methods</i>
<i>EN ISO 62366-1</i>	<i>Medical devices - Part 1: Application of usability engineering to medical devices</i>
<i>EN ISO 14971</i>	<i>Medical devices - Application of risk management to medical devices</i>

Authorised Signatory**Name-Surname :** Murat YILDIZ**Position :** CEO**Signed :**

**MY TİCARETVE
MEDİKAL ANONİM ŞİRKETİ**
Ömerli Mah. General Sükrü Koraltı Cad.
No: 33 Arslanbey / İSTANBUL
Büyükdere V.D.: 626 040 4605
Tel: 0212 332 29 64 Fax: 0212 438 20 65
www.mymedikal.com.tr

Dated : 22.03.2020

TECHNICAL FILE

MUMU SURGICAL MASK



MY TİCARET VE MEDİKAL A.Ş.
Ömerli Mahallesi General Şükrü Koraltı Caddesi No:33 Arnavutköy İstanbul/TURKEY
www.mymedikal.com.tr info@mymedikal.com.tr

TECHNICAL FILE
Mumu 3-Ply Surgical Mask
(Facial Mask for Medical Use)

0. DEFINITION :

Mumu 3-ply surgical mask (facial mask for medical use), can be fit according to each face measures and shapes, flexible and can be used without disturbing the soft structure. Air permeable and lets breathing easily. Non-irritating. Provides protection against bacteria.

1. QUALITY SYSTEM OF MY TICARET VE MEDIKAL :


MY TICARET VE MEDIKAL has been manufacturing its products with the quality systems as given below;

- EN ISO 13485:2016 Quality Assurance System-Medical Devices
- EN ISO 9001:2015+AC Quality Assurance System

2. PRODUCT IDENTIFICATION AND RECOMMENDED USE :

The product is made by non woven fabric. The product's composition is polypropylene and does not includes latex. The product is breathable and has no special personal on environmental hazards. The product is made automatically in hygienic conditions. The product prevents the potential reactions between all kind of liquids and particles, microorganisms.

Ref No:	MM.NS.LM.01
----------------	-------------



Dimension	Body Size	Length	175 mm
		Width	95 mm
	Height of Earloop		70 mm
	Length of Earloop		160 mm
	Pleat Depth		10 mm
	Length of Nose-piece		90 mm

TECHNICAL FILE
Mumu 3-Ply Surgical Mask
(Facial Mask for Medical Use)

Characteristic	Specification		
Materials	Mask Body	Outer Material	Spunbond 25 -30 gr
		Filter Layer	Meltblown 25 gr
		Inner Material	Spunbond 20-25 gr

3. SUBSTANCE/MIXTURE OF RAW MATERIAL

3.1. Material Safety Data Sheet (MSDS)

3.1.1. Composition:

Identification of the type of nonwoven product.

Thermobonded Nonwoven by calendering process

Polypropylene CAS No: 25085-53-4 spunbond–meltblown modification

Web surface treatment - Concentration above 1%0 *No.

Binder..... *No

Additives Yes.

May contain color pigments if the product is in color. White color may contain TiO2.

Other major components *No

Chemicals (in relevant concentration) that are in list of dangerous substances.....*.No

3.1.2. Hazards identification

Under normal conditions of use and handling, this product is not expected to create any health or safety hazards.

Accidental thermal decomposition or melting state can present hazards.

3.1.3. First-aid measures

Undernormal condition ;

Inhalation	No specific measure to be taken
Skin contact.....	No specific measure to be taken
Eyes contact	No specific measure to be taken
Ingestion.....	No specific measure to be taken

3.1.4. Fire fighting measures

1. Suitable extinguishing media;

Water spray, dry chemical or CO2 extinguisher. No special procedures are expected to be necessary for this product. Normal fire fighting procedures should be followed to avoid inhalation of smoke and gases

2. Extinguishing media not to be used..... None

3. Special exposure hazards..... For flammable and toxic fumes as well as skin contact with molten materials see § 10

4. Special protective clothing for fire-fighter None. It is recommended that fire-



TECHNICAL FILE
Mumu 3-Ply Surgical Mask
(Facial Mask for Medical Use)

fighters should wear full protective clothing including self contained breathing apparatus.

3.1.5. Accidental release measures

Personal Precautions: Avoid dust formation. Forms slippery surface.

3.1.6. Handling and storage

Keep in a dry and closed area with the original packing. The packages have to be handled so that they can not break and to be arranged as to prevent them from falling. The goods shall be handled with good industrial hygiene and safety practice.

3.1.7. Exposure controls / personal protection

No specific measures. Handle in accordance with good industrial hygiene and safety practice. Use of safety glasses and face mask is recommended if dust is formed during application.

3.1.8. Physical and chemical properties

Aspect	Solid, in rolls or sheets
Appearance (the colour of the product as supplied)	Normally white if not a specific color is mentioned.
Odour	Practically odorless
Ph	Not applicable
Boiling point/boiling range	Not applicable
Melting point/melting range.....	(Polypropylene 165°C (330°F))
Decomposition temperature	> 260°C (500°F)
Flash point.....	Not applicable
Flammability	Not easily flammable
Explosive properties.....	Not applicable
Oxidizing properties.....	Not applicable
Vapourpressure	Not applicable
Static electricity	The product can develop and/or accumulate static electricity, (i.e. by rubbing or friction)
Solubility.....	Mater insoluble - fat insoluble
Partition coefficient.....	Not applicable

3.1.9. Stability and reactivity

The product is stable at room temperatures and does not decompose or self react when handled and stored under prescribed conditions. Toxic fumes can be generated under thermal decomposition.

3.1.10. Toxicological information

No toxic reaction known under normal conditions. Particularly, no case of coetaneous sensitisation or of mutagenic / carcinogenic activity is known. Under decomposition conditions, toxic fumes and contaminated water.

3.1.11. Ecological information

For transportation, storage and normal use no toxicological effect known. The fabric will not degrade biologically in short term.

MY TİCARET VE MEDİKAL A.Ş.

Ömerli Mah. General Şükrü Koraltı Cad. No:33 Arnavutköy, İstanbul/TURKEY
Pbx: +90 212 438 20 64 Fax:+90 212 438 20 65 www.mymedikal.com.tr info@mymedikal.com.tr

TECHNICAL FILE
Mumu 3-Ply Surgical Mask
(Facial Mask for Medical Use)

3.1.12. Disposal considerations

As non hazardous solid waste, depending on local registration, nonwovens can be disposed of through recycling, landfill.

3.1.13. Transport information

Not classified as dangerous for transport.

3.1.14. Regulatory information

Not classified as dangerous in compliance with Turkey and European regulation regarding classifying, packaging and labeling of hazardous substances and products.

3.2. Technical Data Sheet and Certificate

TECHNICAL DATA SHEET			
Product		: Polypropylene	
Product Description		: ENDLESS FILAMENTS SPUNBOND, THERMALLY BONDED.	
Raw Material		: 100 % PP	
Application on Fabric		: SB HYDROPHOBIC	
Treatment		:	
Fabric Colour		: WHITE	
Customer Name		:	
Weight		: 25 GSM	
Width		:	
Packing		: PE BAG WITH LABEL	
PROPERTIES	TEST METHOD	UNIT	TARGET
WEIGHT	NWSP 130.1.R3 (15)	gsm	25
THICKNESS	NWSP 120.1.R3 (15)	mm	0,28
TENSILE STRENGTH	MD NWSP 110.4.R3 (15)	N/5 cm	55,0
			CD 25,0
ELONGATION AT BREAK	MD NWSP 110.4.R3 (15)	%	118,0
			CD 114,5
Tolerances For The Average Results			
Weight	± 5 %	Roll Tolerance Length : - 0 / +3% against target / ordered length Width : Up to 150 cm in width = -0mm/+5mm Over 150 cm in width = - 0mm/+10 Splice : Maximum five splices per roll.	
Thickness	± 10 %		
Tensile Strength	± 15 %		
Elongation	± 15 %		
Hydrostatic Head	± 15 %		
Liquid Strike-Through Time	± 0,5 %		
Air Permeability	± 20 %		
Absorption	± 20 %		
The product is wound onto cardboard cores and then wrapped in polyethylene film. Bar code labels with product code, description and lot details are applied to the outside of each pack and a small label is applied to each roll. Suitable sized rolls may be palletised and the pallet load cling-wrapped			
Preparation Date	QUALITY CONTROL APPROVAL		
01.04.2020			

TECHNICAL FILE
Mumu 3-Ply Surgical Mask
(Facial Mask for Medical Use)

Hohenstein Textile Testing Institute GmbH & Co. KG
Schöneweide 1, 74367 Bönnigheim, Germany

OEKO-TEX®
INSPIRING CONFIDENCE

CERTIFICATE

The company

BAYTEKS TEKNIK TEKSTİL SAN. VE TİC. A.Ş.
Organize Sanayi Bölgesi 19 Nolu Cadde No. 9
79000 Merkez - Kilis, TURKEY

is granted authorisation according to STANDARD 100 by OEKO-TEX® to use
the STANDARD 100 by OEKO-TEX® mark, based on our test report
19.0.01029



for the following articles:

Nonwoven spunbonded, melt blown and their composite structures produced from white and masterbatch (pigment dyestuff) dyed polypropylene and polypropylene/polyethylene (reprocessing of own waste), with and without PE lamination (in colour white and blue) and additives including UV stabiliser.

The results of the inspection made according to STANDARD 100 by OEKO-TEX®, Appendix 4, product class II have shown that the above mentioned goods meet the human-ecological requirements of the STANDARD 100 by OEKO-TEX® presently established in Appendix 4 for products with direct contact to skin.

The certified articles fulfil requirements of Annex XVII of REACH (incl. the use of azo colourants, nickel release, etc.), the American requirement regarding total content of lead in children's articles (CPSIA, with the exception of accessories made from glass) and of the Chinese standard GB 18401:2010 (labelling requirements were not verified).

The holder of the certificate, who has issued a conformity declaration according to ISO 17050-1, is under an obligation to use the STANDARD 100 by OEKO-TEX® mark only in conjunction with products that conform with the sample initially tested. The conformity is verified by audits.

The certificate 06.MO.41419 is valid until 31.03.2021

Bönnigheim, 28.01.2020



Dipl.-Ing. (FH) Ivonne Schramm
Head of Certification Body OEKO-TEX®

MY TİCARET VE MEDİKAL A.Ş.

Ömerli Mah. General Şükrü Koraltı Cad. No:33 Arnavutköy, İstanbul/TURKEY
Pbx: +90 212 438 20 64 Fax: +90 212 438 20 65 www.mymedikal.com.tr info@mymedikal.com.tr

TECHNICAL FILE
Mumu 3-Ply Surgical Mask
(Facial Mask for Medical Use)

4. PERFORMANCE REQUIREMENTS OF FINISHED PRODUCTS

We declare our products comply with EN 14683:2019+AC:2019 as TYPE IIR

Parameters	Units	Method	Results	Test Results
BFE (Filtration)	%	Internal method based on EN 14683:2019+AC:2019	≥98	% 98,97 (Eurolab-2020170631)
Differential pressure (Breathability)	Pa/cm ²	Internal method based on EN 14683:2019+AC:2019	<40	21 (Eurolab-2020170631)
Microbial cleanliness / bioburden	UFC/g	ISO 11737-1:2018	≤30	21 (Eurolab-2020170631)
Splash resistance pressure	kPa	ISO 22609	≥16	18 (Eurolab-2020170631)
Biocompatibility		ISO 10993-1	Suitable for skin Cytotoxicity, Irritation, Sensitization	Cytotoxicity: It is not cytotoxicity. (Oxigen-2020-C-1099)

5. INSTRUCTIONS FOR USE

5.1.Intended Use:

Surgical mask (facial mask for medical use) is intended to be worn by medical personnel during surgical or other medical procedures to protect both the patient and the operating personnel and any other person that want and need to be protected, from transfer of microorganisms, body fluid, particulate material transfer and any other microbes.

Reduces exposure to blood and body fluids. Minimizes contamination to exhaled microorganisms. This product is intended for use in infection control practices.

5.2. Technical Specifications:

Non-irritating, Fluid Resistant, Three Ply construction.

3 pleats of folds to allow the user to expand the mask so it covers the area from the nose to the chin. Mask is secured with an ear loop to be placed behind the ears.

MY TİCARET VE MEDİKAL A.Ş.

Ömerli Mah. General Şükrü Koraltı Cad. No:33 Arnavutköy, İstanbul/TURKEY
Pbx: +90 212 438 20 64 Fax:+90 212 438 20 65 www.mymedikal.com.tr info@mymedikal.com.tr

TECHNICAL FILE
Mumu 3-Ply Surgical Mask
(Facial Mask for Medical Use)

The surgical mask's (facial mask for medical use) three-ply layers work as follows:

- **The outer layer** repels water, blood, and other body fluids.
- **The middle layer** filters certain pathogens.
- **The inner layer** absorbs moisture and sweat from exhaled air.

5.3. Donning The Mask:

- Before putting on the mask, wash your hands for at least 20 seconds with soap and water, or rub your hands together thoroughly with alcohol-based hand sanitizer.
- Check for defects in the face mask, such as tears or broken loops.
- Position the colored side of the mask outward.
- If present, make sure the metallic strip is at the top of the mask and positioned against the bridge of your nose.
- Ear loops: Hold the mask by both ear loops and place one loop over each ear.
- Mold the bendable metallic upper strip to the shape of your nose by pinching and pressing down on it with your fingers.
- Pull the bottom of the mask over your mouth and chin.
- Be sure the mask fits snugly.
- Don't touch the mask once in position.
- If the mask gets soiled or damp, replace it with a new one.

!! Do not:

- touch the mask once it's secured on your face, as it might have pathogens on it
- dangle the mask from one ear
- hang the mask around your neck
- reuse single-use masks.

If you have to touch the face mask while you're wearing it, wash your hands first. Be sure to also wash your hands afterward, or use hand sanitizer.

5.4. Doffing The Mask:

MY TİCARET VE MEDİKAL A.Ş.

Ömerli Mah. General Şükrü Koraltı Cad. No:33 Arnavutköy, İstanbul/TURKEY
Pbx: +90 212 438 20 64 Fax:+90 212 438 20 65 www.mymedikal.com.tr info@mymedikal.com.tr

TECHNICAL FILE
Mumu 3-Ply Surgical Mask
(Facial Mask for Medical Use)

- Before you take off the mask, wash your hands well or use hand sanitizer.
- Avoid touching the mask itself, as it could be contaminated. Hold it by the loops, ties, or bands only.
- Carefully remove the mask from your face once you: unhook both ear loops.
- Holding the mask loops discard the mask by placing it in a covered trash bin.
- After removing the mask, wash your hands thoroughly or use hand sanitizer.

6. FIRE FIGHTING MEASURES :

Suitable Fire Extinguishers and Methods:

Water spray, foam, carbon dioxide or dry chemicals. A sudden intervention should be made to the fire exit without any possible danger. If the material is melted, do not apply direct water flow. Use fine water spray or foam.

7. DISPOSAL CONSIDERATIONS :











Dispose of according to the Regulation on Control of Hazardous Wastes.

8. USE AREAS :

Hospitals, medical companies, doctor offices, laboratories, food manufacturers, cleaning companies and work places where the hygienic areas are necessary.

9. SHELF LIFE : 5 Years

10. SYMBOLS :

	Production Date		Shows the conformity to the European standards for the self-declared Class I products
	Manufacturer		Keep Dry
	Expiration Date		Keep out of sunlight
	Lot Number		Non-Sterile
	Referance Number		Do not use again

11. STORAGE :

MY TİCARET VE MEDİKAL A.Ş.

Ömerli Mah. General Şükrü Koraltı Cad. No:33 Arnavutköy, İstanbul/TURKEY
Pbx: +90 212 438 20 64 Fax:+90 212 438 20 65 www.mymedikal.com.tr info@mymedikal.com.tr

TECHNICAL FILE
Mumu 3-Ply Surgical Mask
(Facial Mask for Medical Use)

Capable of being stored continuously in ambient temperature of 10 to 30 deg C and relative

humidity of 15 to 55%. Capable of operating continuously in ambient temperature of 10 to 30 deg C and relative humidity of 15 to 55%.

It is recommended that surgical face masks (facial mask for medical use)should be stored in their original containers and should be stored away from direct sunlight, heat sources and liquids, including chemicals. The area should be clean and should protect the masks from contamination. Never store it in a purse or pocket.

12. SAFETY INFORMATION :

- Pay attention to the warnings.
- It is not sterile.
- It is for single use only.
- Do not use if the package is damaged.
- Do not use the product after expiration date.



13. PACKAGING

Ref. No.

MM.NS.LM.01

Quantity per Box

50 pcs



Quality Assurance Manager

Vedat ÇETİN

MY TİCARET VE
MEDİKAL ANONİM ŞİRKETİ
Ömerli Mah. General Şükrü Koraltı Cad. No:33 Arnavutköy, İstanbul
Büyükdere V.D. : 626 040 4605
Tel: 0212 438 20 64 Fax: 0212 438 20 65
www.mymedikal.com.tr

MY TİCARET VE MEDİKAL A.Ş.

Ömerli Mah. General Şükrü Koraltı Cad. No:33 Arnavutköy, İstanbul/TURKEY
Pbx: +90 212 438 20 64 Fax:+90 212 438 20 65 www.mymedikal.com.tr info@mymedikal.com.tr

5190243IB02

2020170631



Report No: 2020170631
Applicant: MY TİCARET VE MEDİKAL A.Ş.
Ömerli mah. General Şükrü Koraltı cad. No:33 Arnavutköy/ İstanbul
Contact Person: Z. Melek ÖZ BOLAT
Contact Telephone: 0212 438 2064
Contact e-mail: info@mymedikal.com.tr / kalite@mymedikal.com.tr
Sample Accepted on: 10.06.2020
Report Date: 17.06.2020
Total number of pages: 9 (Pg)
Sample ID: Surgical Mask

TEST	METHOD	Specimen	RESULT
* Medical and surgical face masks - Requirements and test methods	EN 14683+AC 2019	Surgical Mask	PASS
			TYPE IIR



Seal

Customer Representative
Hasan KUTLULaboratory Manager
Hava SARIAYDIN

EUROLAB[®] (TÜRCERT TEKNİK KONTROL VE BELGELENDİRME A.Ş.)

It is prohibited to change any and all versions of this document in any manner whatsoever. In case of a conflict between the electronic version (e.g. PDF file) and the original paper version provided by EUROLAB[®], the latter will prevail.

TÜRCERT Teknik Kontrol ve Belgelendirme A.Ş. disclaim liability for any direct, indirect, consequential or incidental damages that may result from the use of the information or data, or from the inability to use the information or data contained in this document.

The contents of this report may only be transmitted to third parties in its entirety and provided with the copyright notice, prohibition to change, electronic versions' validity notice and disclaimer.

Environment

The requirements and standards apply to equipment intended for use in

X	Residential (domestic) environment
X	Commercial and light-industrial environment
X	Industrial environment
X	Medical environment

Requirements and test methods

This European Standard specifies construction, design, performance requirements and test methods for medical face masks intended to limit the transmission of infective agents from staff to patients during surgical procedures and other medical settings with similar requirements. A medical face mask with an appropriate microbial barrier can also be effective in reducing the emission of infective agents from the nose and mouth of an asymptomatic carrier or a patient with clinical symptoms.

General

All tests shall be carried out on finished products or samples cut from finished products, if applicable in their sterile state.

Method for in-vitro determination of bacterial filtration efficiency (BFE)

Principle

A specimen of the mask material is clamped between a six-stage cascade impactor and an aerosol chamber. An aerosol of *Staphylococcus aureus* is introduced into the aerosol chamber and drawn through the mask material and the impactor under vacuum. The bacterial filtration efficiency of the mask is given by the number of colony forming units passing through the medical face mask material expressed as a percentage of the number of colony forming units present in the challenge aerosol.

Reagents and materials

General

Describe commercially available solutions of tryptic soy agar and tryptic soy broth. Other variants may be suitable.

Tryptic soy agar

Formula/liter:

Enzymatic digest of casein	15 g
Enzymatic digest of soybean meal	5 g
Sodium chloride	5 g
Agar	15 g
Final pH	7,3 ± 0,2 at 25 °C

Tryptic soy broth

Formula/liter:

Enzymatic digest of casein	17 g
Enzymatic digest of soybean meal	3 g
Sodium chloride	5 g
Dextrose	2,5 g
Final pH	7,3 ± 0,2 at 25 °C

Peptone Water

Formula/liter:

Peptone	1 g
Sodium chloride	5 g
Final pH	7,3 ± 0,2 at 25 °C

Preparation of bacterial challenge

Staphylococcus aureus shall be inoculated into 30 ml tryptic soy broth in an Erlenmeyer flask and incubated with mild shaking at a temperature of (37 ± 2) °C for (24 ± 2) h. The culture shall then be diluted in peptone water to give a concentration of approximately 5 × 10⁵ cfu/ml.

The bacterial challenge shall be maintained at (2 200 ± 500) cfu per test. The bacterial challenge shall be determined on the basis of experience and previous positive control plates (see B.6.3) and the dilution of the challenge suspension adjusted accordingly. The mean particle size in the bacterial challenge shall be maintained at (3,0 ± 0,3) µm (see B.6.9).

Procedure

Assemble the apparatus in accordance with the flow chart shown in Figure B.1.

Deliver the bacterial challenge to the nebulizer using the peristaltic or syringe pump.

Perform a positive control run without a test specimen. Initiate the bacterial challenge by turning on the vacuum pump and adjust the flow rate through the cascade impactor to 28,3 l/min. Deliver the bacterial challenge for 1 min. Maintain the airflow through the impactor for 2 min. Then remove the plates from the impactor. Ensure that each plate is numbered to indicate its position in the impactor.

Place fresh plates in the impactor, fix a test specimen in place and repeat the above procedure.

Repeat this procedure for each test specimen.

After the last test specimen has been tested, perform a further positive control run.

Perform a negative control run by passing air, without addition of the bacterial challenge, through the cascade impactor for 2 min.

Incubate all the plates at (37 ± 2) °C for (48 ± 4) h.

For each specimen and control run, count the number of colonies on each plate and add up the counts to give the total number of cfu collected by the impactor using the “positive hole” conversion table1) in accordance with the instructions of the cascade impactor manufacturer. For the two positive control runs, take the mean of the two totals. From the positive control plates calculate the mean particle size of the bacterial challenge aerosol using the “positive hole” conversion table in accordance with the instructions of the cascade impactor manufacturer.

Calculation of bacterial filtration efficiency

For each test specimen calculate the bacterial filtration efficiency B, as a percentage, using the following formula:

$$B = (C - T) / C \times 100$$

Where;


C is the mean of the total plate counts for the two positive control runs;

T is the total plate count for the test specimen.

Method for determination of breathability (differential pressure)

Principle

A device which measures the differential pressure required to draw air through a measured surface area at a constant air flow rate is used to measure the air exchange pressure of the medical face mask material, as shown in Figure 1. Water-filled manometers (M1 and M2) are used to measure the differential pressure. A flow meter is used for measurement of the airflow. An electric vacuum pump draws air through the apparatus and a needle valve is used to adjust the airflow rate.



- Key**
- 1 air inlet
 - 2 flow meter
 - 3 manometer M1
 - 4 filter material
 - 5 manometer M2
 - 6 valve
 - 7 vacuum pump

Figure 1 — Apparatus for measuring air resistance

Procedure

The test specimen is placed across the 2,5 cm diameter orifice (total area 4,9 cm²) and clamped into place so as to minimise air leaks and that the tested area of the specimen will be in line and across the flow of air.

The pump is started and the flow of air adjusted to 8 l/min.

The manometers M1 and M2 are read and recorded.

The procedure described in steps 1 through 3 is carried out on 5 (or appropriate number of) different areas of the mask and the readings averaged.

Calculation of differential pressure

For each test specimen calculate the differential pressure ΔP as follows:

$$\Delta P = (X_{m1} - X_{m2})/4,9$$

Where;

X_{m1} is pressure in Pa, manometer M1, mean of 5 test areas, low pressure side of the material;

X_{m2} is pressure in Pa, manometer M2, mean of 5 test areas, high pressure side of the material;

4,9 is the cm² area of the test material;

ΔP is the differential pressure per cm² of test material expressed in Pa.

Splash resistance

When tested in accordance with ISO 22609 the resistance of the medical face mask to penetration of splashes of liquid shall conform to the minimum value given for Type IIR in Table 1.

Microbial cleanliness (Bioburden)

When tested according to EN ISO 11737-1 the bioburden of the medical mask shall be ≤ 30 cfu/g tested (see Table 1).

To determine the mask's bioburden according to EN ISO 11737-1, follow the procedure below:

The number of masks that shall be tested is minimum 5 (five), but can be greater if necessary to allow for an AQL of 4 %.

Weigh each mask prior testing. The full mask is aseptically removed from the packaging and placed in a sterile 500 ml bottle containing 300 ml of extraction liquid (1 g/l Peptone, 5 g/l NaCl & 2 g/l polysorbate surfactant 20 [e.g. Tween 20, Alkest TW 20]).

The bottle is laid down on an orbital shaker and shaken for 5 min at 250 rpm. After this extraction step, 100 ml of the extraction liquid is filtered through a 0,45 μ filter and laid down on a TSA plate for the total viable aerobic microbial count. Another 100 ml aliquot of the same extraction liquid is filtered in the same way and the filter plated on Sabouraud Dextrose agar (SDA) with chloramphenicol for fungi enumeration. The plates are incubated for 3 days at 30 °C and 7 days at (20 – 25) °C for TSA and SDA plates respectively.

The total bioburden is expressed by addition of the TSA and SDA counts.

In the report, indicate the total bioburden per mask and based on the mask weigh, the total bioburden per gram tested.

TEST REQUIREMENTS

Test	Type I ^a	Type II	Type IIR
Bacterial filtration efficiency (BFE), (%)	≥ 95	≥ 98	≥ 98
Differential pressure (Pa/cm ²)	< 40	< 40	< 60
Splash resistance pressure (kPa)	Not required	Not required	≥ 16,0
Microbial cleanliness (cfu/g)	≤ 30	≤ 30	≤ 30



^a Type I medical face masks should only be used for patients and other persons to reduce the risk of spread of infections particularly in epidemic or pandemic situations. Type I masks are not intended for use by healthcare professionals in an operating room or in other medical settings with similar requirements.

TEST RESULTS**EN 14683 Inspection****SAMPLE : SURGICAL MASK**

Test	Type			Result	Evaluation	
Bacterial filtration efficiency (BFE), (%)	1 ≥ 95	2 ≥98	3 ≥98	98,86	98.97	PASS
				99,01		
				98,92		
				99,03		
				99,04		
Differential pressure (Pa/cm ²)	< 40	<40	<60	21	PASS	
Splash resistance pressure (kPa)	N/A	N/A	≥16,0	18	Type IIR	
Microbial cleanliness (cfu/g)	≤ 30			21	PASS	

Free Area

MASK IMAGES UNDER TEST



*****End of Report*****

TÜRKAK
TÜRK AKREDİTASYON KURUMU
Tarafından Akredite Edilmiştir.
MUAYENE VE ANALİZ RAPORU

AB-0953-T
2020-C-1099
06-2020

Report Number : 2020-C-1099 Date of Report : 05/06/2020
Purpose of Analysis : Cytotoxicity Test
Customer name/address : MY Ticaret ve Medikal. A.S. / Ömerli Mah. Genral Sükrü Koraltı Cad. No:33 / İSTANBUL
Name and identity of test item : Surgical Mask
Code of Sample : Lot: SD20200310
Package of Sample/Quantity : 3 piece
Date of receipt of test item : 28/05/2020
Date of Test/End of test : 29/05/2020 - 05/06/2020
Number of pages : 5

Analysis	Unit	Result	Limit Of Measurement	Recovery	Uncertainty of Meas.	Analysis Metod	Com.
1-*InvitroCytotoxicity Test		it is not Cytotoxicity				TS EN ISO 10993-5(Biological evaluation in medical devices Part 5: Test for in vitro cytotoxicity) TS EN ISO 10993-12 (Biological evaluation in medical devices Part 12: Test sample preparation and Reference Materials..	A

Explanation:

1. Experiment environment

CELL LINE:L929 (Mouse Fibroblast cell)

Culture Medium : DMEM+ L-Glutamin
Fetal Bovine Serum

Penisilin- Streptomisin

Blank :Sterile cell culture medium

NEGATIVE CONTROL:Polietilen Kryo Tüp + Cell

POSITIVE CONTROL:Natural Rubber Latex+ Cell

Report Number

: 2020-C-1099

Date of Report

:05/06/2020

2.METHOD OF APPLICATION

Extraction was performed according to TS EN ISO 10993-12 standard. The samples were placed in a waterbath at a rate of 50 rpm at 37°C for 24 hours in a 10% serum-containing cellculture medium of the size specified in the standard. The extraction was then terminated and the extract obtained was used within 24 hours.

3.ANALYSIS METHOD

Qualitative Evaluation:

Cells were expected to become confluent by sowing 6 well plates.

Subsequently, the 37°C 5% CO2 sample was exposed to negative, positive control and sample extracts for 24 hours. After incubation, cells were microscopically examined and evaluated according to TS EN ISO 10993-5 standard.

Quantitative Evaluation:

In the study, it was applied according to the "TS EN ISO 10993-5 / XTT Cytotoxicity Experiment" standard. The 96-well plate was counted as 100 / well and the cultured cells were incubated for 24 hours to provide 80% confluency. Subsequently, the cells were exposed to 1/1 - dilutions of the sample extract for 4 hours.

At the end of the process, 1 mg / mL XTT was added to the wells and the plates were incubated for 3 hours at 37 ° C in 5% CO2. The assay was terminated by the addition of isopropyl alcohol to the wells and the % viability values were calculated by measuring the color change in the plates (570-650 nm) spectrophotometer.

4. TEST RESULTS

Qualitative Evaluation:

The qualitative evaluation was made according to Table 1 in TS EN ISO 10993-5 standard.

Must No.	Test Material	Reaction	Situations of Cultures
1	Negative Control	0	Discreteendoluminalgranules, celldisruptionno, nodecrease in cellproliferation
2	Positive Control	4	Nearlyallcelllayers have been destroyed
3	Sample	0	Discreteintraoplasmagranules, no cell destruction, no decrease in cell proliferation

Quantitative Evaluation:

(TS EN ISO 10993-5 / XTT Cytotoxicity Test)

Table 2. XTT Test results

DILUTION RATIOS						
TEST NUMBER	100%	75%	50%	25%		
1. AGAIN	0,963	1,111	1,245	1,345		
2. AGAIN	0,914	1,120	1,216	1,337		
3. AGAIN	0,946	1,114	1,224	1,359		
AVERAGE	0,941	1,115	1,228	1,337		
POSITIVE CONTROL						
	100%	75%	50%	25%		
1. AGAIN	0,104	0,206	0,321	0,426		
2. AGAIN	0,106	0,208	0,314	0,441		
3. AGAIN	0,108	0,201	0,325	0,405		
AVERAGE	0,106	0,205	0,320	0,424		
Negative Control(%100)						
	1.Again	2.Again	3.Again			
%100 Ekstrakt	1,109	1,111	1,112			
AVERAGE	1,11					
Blank	A2	A3	A4	A5	A6	A7
	0,888	0,990	0,999	0,996	1,010	1,002
	H2	H3	H4	H5	H6	H7
	0,991	0,992	0,994	0,999	1,080	1,099
AVERAGE	1,003					

Viab.%=100 X OD450e/OD450b

OD450e : % 100 optical density of the sample extract

OD450b : Average value of optical density of blank

Report Number

: 2020-C-1099

Date of Report

:05/06/2020

Test Sample Viab.% : % 94

Positive Control Viab.% : % 11

Negative Control Viab.% : % 111

REVIEWS :

1.The test was carried out in accordance with the standard "TS EN ISO 10993-5 Biologicalevaluation of medicaldevices-Part 5: extracorporealcytotoxicitytests".

2.The effect of the extracts on the cells for qualitative evaluation was examined microscopically andevaluatedbythequalitativemorphologicalgrading of thecytotoxicity of theextractsgiven in the standard "Table 1. Accordingly, the negative control showed no toxic effect on the cells (0), and the positive control showed toxicity as high as expected (4). Since thecytotoxiceffect of thesampleextractswas not toxicwhenexamined, it wasevaluated as (0).Accordingtothestandardused, as indicated in table 1, the presence of a larger rating value of (2) is considered a cytotoxiceffect.

3.The "TS EN ISO 10993-5 / XTT Cytotoxicity Experiment" wasused as the quantitative evaluation method and the obtained results (Table 2) were evaluated statistically. Results from the negative and positive controls used and test validity criteria are met.

In this experiment, the effects of 1/1 dilutions of sample extract on cells were examined; The complete dilution of extractfromthesample (1/1) andviability was **94%**.

According to the standard used, this value is lessthan 70%, indicatingthatthere is nocytotoxiceffect on thesampleextracts since there is a cytotoxicityindicator.

Report Number

: 2020-C-1099

Date of Report

:05/06/2020

Chart1. Qualitative morphological grading of cytotoxicity of extracts

Degree	Reaction	Situations of Cultures
0	No	Discrete intraoplasma granules, no cell destruction, no decrease in cell proliferation
1	Very little	There are more than 20% of cells that are not round, poorly adherent, and contain few or no intracellular granules, or morphologically altered, rarely destroyed cells, only slight growth inhibition can be observed
2	Light	Round cell number is less than 50%, no intraploisio granules, observable cell inhibition is not more than 50%
3	Middle	The number of cells rounded or destroyed is not more than 70%, the cell layers are not completely degraded, the observable cell inhibition is more than 50%
4	Severe	Nearly all cell layers have been destroyed

(*) Analysis method is in scope of accreditation.

Evaluation:

The above mentioned values were determined as the result of the inspection and analysis.

1. No part of this analytical report can be used alone or separately. Unsigned and unsealed reports are defund.

2. Analysis results are valid for the above sample

3. When necessary, "Measurement Uncertainty" and "Recover" information are given together with the analysis results

4. Judicial and administrative procedures to be used for advertising purposes. It can not be partially reproduced and published without permission

Abbreviations: N.A: Not Detected A: Appropriate IA: Inappropriate AF: Assessment Failed EVL: Evaluation

Çel Microbiology Unit Responsible
Havva Lamia Demir (v)

Responsible of the Department of Sample Admission
Nilsun AŞCI

Approved by
05/06/2020
Mehmet Nur ERAT
Laboratory Manager

LABORATUVAR HİZ. TİC. LTD. ŞTİ.
Mah. Hadımköy Bağlantı Yolu Ufuk Plaza
2 No. B. Çekmece-İSTANBUL / TÜRKİYE
Tic. Sic. No: 272 886 85 05 - 08
Kis. Sic. No: 272 886 85 05 - 08



**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**
Esenyurt Firuzköy Bulvarı No:29 34325 Avcılar
İstanbul/ TÜRKİYE

TEST REPORT
DENEY RAPORU



Test
TS EN ISO/IEC 17025
AB-0583-T

AB-0583-T
20018616
06-20

Customer name: MY TİCARET VE MEDİKAL A.Ş.
Address: Ömerli Mahallesi General Şükrü Koraltı Caddesi No:33 ARNAVUTKÖY/
İSTANBUL
Buyer name: -
Contact Person: Z.MELEK ÖZ BOLAT
Order No: -
Article No: -
Name and identity of test item: Beyaz elastik kordon.
The date of receipt of test item: 10.06.2020
Re-submitted/re-confirmation date: -
Date of test: 10.06.2020-17.06.2020
Remarks: -
Sampling: The results given in this report belong to the received sample by vendor.
End-Use: -
Care Label: Not specified.
Number of pages of the report: 3

Türk Akreditasyon Kurumu (TÜRKAK) deney raporlarının tanınması konusunda Avrupa Akreditasyon Birliği (EA) ve Uluslararası Laboratuvar Akreditasyon Birliği (ILAC) ile karşılıklı tanıma antlaşmasını imzalamıştır.

Deney laboratuvarı olarak faaliyet gösteren EKOTEKS LABORATUVAR ve GÖZETİM HİZMETLERİ A.Ş. TÜRKAK'tan AB-0583-T akreditasyon dosya numarası ile ISO 17025:2017 standardına göre akredite edilmiştir.

Deney ve/ veya ölçüm sonuçları, genişletilmiş ölçüm belirsizlikleri (olması halinde) ve deney metodları bu sertifikanın tamamlayıcı kısmı olan takip eden sayfalarda verilmiştir.



Date
18.06.2020

Customer Representative
Özlem ULUS

Head of Testing Laboratory
Sevim A. RAZAK

This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not valid.

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**


AB-0583-T

20018616

06-20

İSTENEN TESTLER	SONUÇ	AÇIKLAMA
FİZİKSEL ÖZELLİK TESTLERİ		
Malzeme Tayini	-	
İstenen değerler müşteri tarafından belirtilmemiştir.		

NOT: Aksi belirtilmediği takdirde testler ile ilgili kayıtlar 5 yıl, orjinal numuneler 3 ay saklanır. Müşteri tarafından talep edildiğinde, testlere ait ölçüm belirsizliği raporlanır fakat "Geçer/Kalır" değerlendirmesinde ölçüm belirsizliği değeri dikkate alınmaz. Raporlanan belirsizlik, genişletilmiş belirsizlik olup standart belirsizlik kapsam faktörü k=2 kullanılarak elde edilmiştir. Güvenilirlik düzeyi % 95'tir. Bu raporda (*) işaretli deneyler akreditasyon kapsamına dahil değildir.



Bu rapor, laboratuvarın yazılı izni olmadan kısmen kopyalanıp çoğaltılamaz.
İmzasız ve mühürsüz raporlar geçersizdir.

AB-0583-T

20018616

06-20

TEST SONUÇLARI

MALZEME TAYİNİ: EKOTEKS 40
FT-IR. Spektrometre test cihazı.

SONUC
Poliüretan

İSTENEN
-

Not: Bu test sonucu FT-IR spektrometre yöntemi ile Poliüretan referans malzemesi ile karşılaştırılarak tespit edilmiştir. Referans test numunesine benzerliği % 73 bulunmuştur.

Not: Lateks içermemektedir.



**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**
Esenyurt Firuzköy Bulvarı No:29 34325 Avcılar
İstanbul/ TÜRKİYE



TEST REPORT
DENEY RAPORU

AB-0583-T

20018616-
ing


06-20

Customer name: MY TİCARET VE MEDİKAL A.Ş.
Address: Ömerli Mahallesi General Şükrü Koralı Caddesi No:33 ARNAVUTKÖY/
İSTANBUL
Buyer name: -
Contact Person: Z.MELEK ÖZ BOLAT
Order No: -
Article No: -
Name and identity of test item: One sample of white elastic cord.
The date of receipt of test item: 10.06.2020
Re-submitted/re-confirmation date: -
Date of test: 10.06.2020-17.06.2020
Remarks: -
Sampling: The results given in this report belong to the received sample by vendor.
End-Use: -
Care Label: Not specified.
Number of pages of the report: 3

The Turkish Accreditation Agency (TÜRKAK) is signatory to the multilateral agreements of the European co-operation for the Accreditation (EA) and of the International Laboratory Accreditation (ILAC) for the Mutual recognition of test reports.

EKOTEKS LABORATUVAR ve GÖZETİM HİZMETLERİ A.Ş. accredited by TÜRKAK under registration number [AB-0583-T] for ISO 17025:2017 as test laboratory.

The test and/or measurement results, the uncertainties (if applicable) with confidence probability and test methods are given on the following pages which are part of this report.



Date
18.06.2020

Customer Representative
Özlem U. U.S.

Head of Testing Laboratory
Sevim A. RAZAK

*This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not valid.*

EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.

AB-0583-T

20018616-
ing

06-20

REQUIRED TESTS	RESULT	COMMENTS
PHYSICAL PROPERTIES TESTS		
Determination of Material	-	
No requirement was given.		

REMARK: Original samples are kept for 3 months and all technical records are kept for 5 years unless otherwise specified. If requested, measurement uncertainty will be reported. But unless otherwise specified, measurement uncertainty is not considered while stating compliance with specification or limit values. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95 %. Tests marked (*) in this report are not included in the accreditation schedule.



This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not valid.

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**

AB-0583-T

20018616-
ing

06-20

TEST RESULTS

DETERMINATION OF MATERIAL: EKOTEKS 40
FT-IR. Spectrophotometer Test Machine.

RESULT
Polyurethane

REQUIREMENT
-

Note: The test result was identified as Polyurethane using by FT-IR spectrometer method which is based on to compare with the reference material. The similarity of the test sample to reference is %73 was found.

Note: Latex does not contains.