



PRODUCT SPECIFICATION

Article name Clungene Covid-19 Antigen rapid test for lay use

Manufacturer Hangzhou Clongene Biotech Co., Ltd.

Anterior nasal procedure

Smear in the nasal vestibule at a depth of about 1.5 cm in the nose

Test result readable after 15 minutes

Sensitivity 97.1% (145 of 148 positive)

Specificity 99.4% (462 of 465 negative)

Cross reactivity none

BfArM special approval lay approval (5640-S-168/21) issued on March 16, 2021

Shelf life / storage 18 months at 4 - 30 ° C

Protect from direct sunlight

Package content 5 x test cassette

5 x sampling tube

5 x extraction buffer

5 x sterile swabs

German instructions for use

Packing 5 packs

190 x 70 x 38 mm

Package information 240 packs / box = 1,200 tests / box, 590 x 400 x 580 mm, approx.
20 kg

